

No. 7-1, Jhongsing Road, Tucheng City, Taipei County, 236, Taiwan T (886) 2 2268-0389 F (886) 2 2268-0639 www.gwinstek.com

GPP-1205 & GPP-1323 Specification

The specifications apply when the GPP-1205/1323 are powered on for at least 30 minutes under +20 °C to +30 °C.









European Jack

European Jack



Rear Panel (including optional GPIB)

CRECIFICATIONS			
SPECIFICATIONS		CDD 1222	CDD 1205
		GPP-1323	GPP-1205
OUTPUT RATING			
Output Voltage		0.000 V to 32.000 V	0.000 V to 20.000 V
Output Current		0.0000 A to 3.0000 A	0.0000 A to 5.0000 A
Output Power		96 W	100 W
CONSTANT VOLTAGE OP	ERATION		
Line Regulation		± (0.01 % of setting + 3 mV)	
Load Regulation		\leq 0.01 % + 3 mV (rating current \leq 3 A)	
		\leq 0.02 % + 5 mV (rating current > 3 A)	
Transient Response		< 100 μs	
Ripple Noise		0.8 mVrms	
Setting Range		0 V to 33.600 V	0 V to 21.000 V
Rise Time		≤ 100 ms	
Fall Time		≤ 100 ms	
Maximum Remote Sensing Compensation Voltage (Single Line)		0.5 V	
Temperature Coefficient (TYP.)		300 ppm/°C	
CONSTANT CURRENT OF	PERATION		
Line Regulation		≤ 0.1 % + 3 mA	
Load Regulation		≤ 0.1 % + 3 mA	
Setting Range		0 A to 3.1500 A	0 A to 5.2500 A
Ripple Noise (Arms)		≤ 2 mArms	
Temperature Coefficient (TYP.)		300 ppm/°C	
RESOLUTION			
Voltage	Programming/Readback	1 mV / 0.1 mV	
Current	Programming/Readback	(High) 0.1 mA / 0.01 mA;	

GOOD WILL INSTRUMENT CO., LTD.



No. 7-1, Jhongsing Road, Tucheng City, Taipei County, 236, Taiwan T (886) 2 2268-0389 F (886) 2 2268-0639 www.gwinstek.com

		(84:44-) 70 - 8 / 7 - 8		
		(Middle) 10 μA / 1 μA;		
METER		(Low) 1 μA/ 0.1 μA		
Full Scale	Voltage/Current	33.6000 V / 3.1500 A	21.0000 V / 5.2500 A	
Programming Resolution	Voltage/Current	5 digits / 5 digits	21.0000 V / 3.2300 A	
Readback Resolution	Voltage/Current	6 digits / 6 digits		
Setting Accuracy	voltage/ carrent	Voltage: ± (0.03 % of read	ling + 10 mV)	
		Current: ± (0.3 % of readi	,	
		Current: ± (0.3 % of readi	, , ,	
		Current: ± (0.3 % of readi		
Readback Accuracy		Voltage: ± (0.03 % of reading + 10 mV)		
		Current: ± (0.3 % of reading + 10 mA) (H)		
		Current: ± (0.3 % of readi		
		Current: ± (0.3 % of readi	· , , ,	
DC LOAD MODE			, , ,	
Display	Voltage	3.000 V to 32.000 V	3.000 V to 20.000 V	
- ,	Current	0 A to 3.0000 A	0 A to 5.0000 A	
	Power	96 W	100 W	
CV Mode	Setting Range	3.000 V to 32.000 V	3.000 V to 20.000 V	
CV Wode			3.000 V to 20.000 V	
	Setting/Readback Accuracy	≤ 0.1 % + 30 mV		
	Resolution	1 mV		
CC Mode		0 A to 3.0000 A	0 A to 5.0000 A	
CC Wode	Setting Range		0 A to 3.0000 A	
	Setting/Readback Accuracy	\leq ± 0.3 % + 10 mA		
	Resolution	0.1 mA		
PROTECTION	Resolution	0.1 111/1		
OVP	Setting Range	1.8 V to 35.2 V	1.0 V to 22.0 V	
OVI	Setting Accuracy	± 100 mV	1.0 ¥ to 22.0 ¥	
	Operation	Turns the output off, disp	lavs OVP	
ОСР	Setting Range	0.15 A to 3.3 A	0.25 A to 5.5 A	
	Setting Accuracy	± 20 mA		
	Operation	Turns the output off, disp	lays OCP	
ОТР	Operation		Turns the output off, displays OTP	
Insulation Resistance		Between chassis and terminal: 20 M Ω or above (DC		
		500V)		
			power cord: 30 M Ω or above	
CEDIEC AND DADALLEI C	A DA DU ITO	(DC 500V)		
SERIES AND PARALLEL CA	APABILITY			
Parallel Number		4 units		
Series Number		4 units	4 units	
ADVANCED FUNCTION				
Trigger Signal *1	Trigger Input		S signal is applied for 100 μs	
		or longer. It receives a pulse to perfo	orm actions like newer	
		output, V/I set operation		
	Trigger Output	Trigger output: approx. 3.		
		1ms, Output impedance:		
		It outputs a pulse when p		
		operation or memory reca	all is executed.	
Status Signal Out *1 *2	OUT ON/OFF Status	Turns on when the output		
	CV Status	Turns on during CV opera	tion	

GOOD WILL INSTRUMENT CO., LTD.



No. 7-1, Jhongsing Road, Tucheng City, Taipei County, 236, Taiwan T (886) 2 2268-0389 F (886) 2 2268-0639 www.gwinstek.com

		T =	
	CC Status	Turns on during CC operation	
	ALM Status	Turns on when an alarm has been activated	
	PWR ON Status	Turns on when the power is turned on	
GENERAL SPECIFICATION	DNS		
Display		2.4-inch TFT LCD	
Interface	LAN	MAC Address, Gateway IP Address, Instrument IP	
	LAN	Address, Subnet Mask	
	USB	Type A: Host, Type B: Slave, Speed: 1.1/2.0,	
		USB-CDC/TMC	
	GPIB (Factory Optional)	SCPI-1993, IEEE 488.2 compliant interface	
Operating Environment		Indoor use, Overvoltage Category II, Altitude: ≤ 2000	
		m, Ambient temperature: 0 °C to 40 °C, Relative	
		humidity: 20 % to 80 % RH; No condensation	
Storage Environment		Ambient temperature: -20 °C to 70 °C, Relative	
•		humidity: 20 % to 85 % RH; No condensation	
Power Source		AC (100 V, 120 V, 220 V, 240 V) ± 10 %, 50 or 60 Hz	
Power Consumption		300 VA	
Max. Inrush Current		30 A max or less	
Dimensions & Weight		107 mm x 124 mm x 313 mm (W x H x D) (not	
		including protrusions), Approx. 5.5 kg	

This information is subject to change without notice.

(*1) The specification refers to the highest resolution. (*2) Only channel 1 and channel 2 can perform series and parallel functions.

ORDERING INFORMATION				
GPP-1205	100 W Single Channel Programming Linear DC Power Supply (USB, LAN) (20 V /5 A)			
GPP-1323	96 W Single Channel Programming Linear DC Power Supply (USB, LAN) (32 V/3 A)			
A CCECCODU	-6			
ACCESSORIE				
Power Cord x 1, Packing List x 1, Test lead: Non-European: GTL-104A x 1 Test lead: European:				
GTL-204A x				
OPTION (Ma	anufacturer Installed Only)			
GPIB interfac	ce control of the con			
OPTIONAL A	ACCESSORIES			
GTL-303	RF Cable, for Trigger In/Out use			
GTL-246	USB Cable (USB 2.0 A-B Type, approx. 1200mm)			
GRA-441-J	Rack Mount Kit for JIS type			
GRA-441-E	Rack Mount Kit for EIA type			
FREE DOWN	ILOAD			
PC Software,	LabVIEW Driver			