GOOD WILL INSTRUMENT CO., LTD. N0. 7-1, Jhongsing Road, Tucheng Dist., New Taipei City, 236, Taiwan T (886) 2 2268-0389 F (886)2 2268-0639 www.gwinstek.com

**G<u><u></u></u>INSTEK** 

## **GPS-Series Specifications**

The specifications apply when the GPS-Series is powered on for at least 30 minutes under +20°C~+30°C.

DC OUTPUT	Output Malta (M)	Output America (A)
	Output Volts (V)	Output Amps (A)
GPS-1850	0~18	0~5
GPS-3030	0~30	0~3
GPS-6010	0~60	0~1
GPS-1830D	0~18	0~3
GPS-1850D	0~18	0~5
GPS-3030D	0~30	0~3
GPS-3030DD	0~30	0~3
CONSTANT VOLTAGE (		
Regulation	Line regulation $\leq 0.01\% + 3mV$	
	Load regulation $\leq$ 0.01% + 3mV (rating current $\leq$ 3A) Load regulation $\leq$ 0.01% + 5mV (rating current > 3A)	
Ripple & Noise	$\leq 0.5$ mVrms 5Hz ~ 1MHz (rating current $\leq 3A$ )	
	$\leq 1 \text{mVrms 5Hz} \sim 1 \text{MHz} \text{ (rating current > 3A)}$	
Recovery Time	$\leq 100\mu$ S ( 50% load change, minimum load 0.5A )	
Temp. Coefficient	≦300ppm/°C	
Output range	0 to rating voltage continuously adjustable	
CONSTANT CURRENT OPERATION		
Regulation	Line regulation $\leq$ 0.2% + 3mA	
	Load regulation $\leq$ 0.2% + 3mA	
Ripple Current	≦3mArms	
Output Range	0 to rating current continuously adjustable (Hi/Lo range switchable)	
METER		
Analog	V-meter and I-meter	
	2.5 class	
	Dimensions 50 x 50 mm	
Digital	31/2 digital 0.5" LED display (3030D)	
	31/2 digital 0.39" LED display (3030DD)	
	Accuracy ±(0.5% of rdg + 2 digits)	
INSULATION		
Chassis and Terminal	20MΩ or above (DC 500V)	
Chassis and AC Cord	30MΩ or above (DC 500V)	
POWER SOURCE		
	AC 100V/120V/220V/240V±10%, 50/60Hz	
ACCESSORIES		
	User manual x 1	
	Power cord x 1	
	Test lead GTL-105 x 1	
	European test lead GTL-201A (optional)	
DIMENSIONS & WEIGH	IT	
	128(W) x 145(H) x 285(D)mm	
	Approx. 5kg	